

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00585	Calib. Date: 10/09/2009
Certification File No.:	00586	Cert. No.: 00014
Linearity File No.:	00587	Cert. Date: 10/09/2009
Solution File No.:	00620	Lin. No.: 00010
Sequential File No.:	00620	Lin. Date: 10/09/2009
		Soln. Date: 01/27/2010
		Soln. No.: 00078
		File Date: 01/27/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	09E066	Expires: 05/22/2011
		Bottle No.: 0199

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:49S		
Control 1 EC	0.101%	13:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51S		
Control 2 EC	0.098%	13:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52S		
Control 3 EC	0.099%	13:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: Sgt. T. Tarr

Badge No.: 427

Date: 01/27/2010