



WEST WINDSOR TOWNSHIP POLICE DEPARTMENT

20 Municipal Drive * P.O. Box 38
West Windsor, New Jersey 08550

Attn: Discovery Clerk
(609) 799-1222
FAX (609) 799-6515

Records (609) 799-9282
Det. Bur. (609) 799-8263
Traffic (609) 799-8288
Juvenile (609) 799-3533
Community Policing (609) 799-0452

Discovery Upload to Internet

Date: 10/25/2010

Time: Approx. 1144 Hrs

From: Lt. Robert Garofalo

Number of Pages Including This Page: 2

Message (If Any):

Alcotest New Standard Solution Report

Any questions or concerns should be directed to the discovery clerk at the main police number.



Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00694	Calib. Date: 09/15/2010
Certification File No.:	00695	Calib. No.: 00016
Linearity File No.:	00696	Cert. Date: 09/15/2010
Solution File No.:	00712	Cert. No.: 00012
Sequential File No.:	00712	Lin. Date: 09/15/2010
		Lin. No.: 00012
		Soln. Date: 10/19/2010
		Soln. No.: 00092
		File Date: 10/19/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %: 0.100%		Expires: 05/22/2011
Solution Control Lot: 09E066		Bottle No.: 0188

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:32D		
Control 1 EC	0.099%	13:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		
Control 2 EC	0.099%	13:34D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		
Control 3 EC	0.099%	13:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Seft. A. Tarr

Badge No.: 427

Date: 10/19/2010

DDWAP2-216
AT