

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00585	Calib. Date: 10/09/2009
Certification File No.:	00586	Cert. No.: 00014
Linearity File No.:	00587	Cert. Date: 10/09/2009
Solution File No.:	00598	Lin. No.: 00010
Sequential File No.:	00598	Soln. Date: 11/05/2009
		Soln. No.: 00075
		File Date: 11/05/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	09E066	Expires: 05/22/2011
		Bottle No.: 0197

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:44S		
Control 1 EC	0.100%	13:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.099%	13:46S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		
Control 3 EC	0.099%	13:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: Sgt. P. Tarr #427

Badge No.: 427

Date: 11/05/2009