

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00694	Calib. Date: 09/15/2010
Certification File No.:	00695	Calib. No.: 00016
Linearity File No.:	00696	Cert. Date: 09/15/2010
Solution File No.:	00736	Cert. No.: 00012
Sequential File No.:	00736	Lin. Date: 09/15/2010
		Lin. No.: 00012
		Soln. Date: 12/16/2010
		Soln. No.: 00094
		File Date: 12/16/2010
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	09E066	Expires: 05/22/2011
		Bottle No.: 0200

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:03S		
Control 1 EC	0.102%	13:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05S		
Control 2 EC	0.099%	13:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06S		
Control 3 EC	0.099%	13:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR	First Name: THOMAS	MI: S
Signature: <u>Sgt. D. Tarr #427 DDWAP2-216</u>		Badge No.: 427
		Date: 12/16/2010