

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00841	Calib. Date: 10/26/2011
Certification File No.:	00842	Calib. No.: 00026
Linearity File No.:	00843	Cert. Date: 10/26/2011
Solution File No.:	00857	Cert. No.: 00014
Sequential File No.:	00857	Lin. Date: 10/26/2011
		Lin. No.: 00014
		Soln. Date: 12/02/2011
		Soln. No.: 00112
		File Date: 12/02/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0183

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:29S		
Control 1 EC	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 2 EC	0.097%	11:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32S		
Control 3 EC	0.098%	11:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: EVANS

First Name: LEE

MI: F

Signature: \_\_\_\_\_

*PH Lee #550*

Badge No.: 550

Date: 12/02/2011