

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00481	Calib. Date:	11/25/2008
Certification File No.:	00482	Cert. No.:	00011
Linearity File No.:	00483	Cert. Date:	11/25/2008
Solution File No.:	00492	Lin. Date:	11/25/2008
Sequential File No.:	00492	Lin. No.:	00007
		Soln. Date:	12/24/2008
		Soln. No.:	00061
		File Date:	12/24/2008

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	10/03/2009
Solution Control Lot:	07J048			Bottle No.:	0539

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:24S		
Control 1 EC	0.099%	08:25S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:26S		
Control 2 EC	0.098%	08:26S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:27S		
Control 3 EC	0.098%	08:28S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature probe serial number DDWAP2-216

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature:

Badge No.: 427

Date: 12/24/2008