

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00585	Calib. Date: 10/09/2009
Certification File No.:	00586	Cert. Date: 10/09/2009
Linearity File No.:	00587	Lin. Date: 10/09/2009
Solution File No.:	00603	Soln. Date: 12/03/2009
Sequential File No.:	00603	File Date: 12/03/2009

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 05/22/2011
Solution Control Lot:	09E066		Bottle No.: 0198

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:28S		
Control 1 EC	0.100%	15:29S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:30S		
Control 2 EC	0.097%	15:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:31S		
Control 3 EC	0.098%	15:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

PROBE SERIAL # DDWAP2-216

Changed By:

Last Name: EVANS

First Name: LEE

MI: F

Signature: PH Evans # 550

Badge No.: 550

Date: 12/03/2009