

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0382 |
| Location: | WEST WINDSOR POLICE | |
| Calibration File No.: | 00585 | Calib. Date: 10/09/2009 |
| Certification File No.: | 00586 | Calib. No.: 00014 |
| Linearity File No.: | 00587 | Cert. Date: 10/09/2009 |
| Solution File No.: | 00611 | Cert. No.: 00010 |
| Sequential File No.: | 00611 | Lin. Date: 10/09/2009 |
| | | Lin. No.: 00010 |
| | | Soln. Date: 12/30/2009 |
| | | Soln. No.: 00077 |
| | | File Date: 12/30/2009 |

| | | | |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDWJ S3-0363 |
| Control Solution %: | 0.100% | | Expires: 05/22/2011 |
| Solution Control Lot: | 09E066 | | Bottle No.: 0193 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 13:31S | | |
| Control 1 EC | 0.100% | 13:32S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 13:32S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:32S | | |
| Control 2 EC | 0.098% | 13:33S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 13:33S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:34S | | |
| Control 3 EC | 0.099% | 13:34S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 13:34S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:35S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 12/30/2009