

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARWF-0382 |
| Location: | WEST WINDSOR POLICE | |
| Calibration File No.: | 00585 | Calib. Date: 10/09/2009 |
| Certification File No.: | 00586 | Calib. No.: 00014 |
| Linearity File No.: | 00587 | Cert. Date: 10/09/2009 |
| Solution File No.: | 00623 | Cert. No.: 00010 |
| Sequential File No.: | 00623 | Lin. Date: 10/09/2009 |
| | | Lin. No.: 00010 |
| | | Soln. Date: 02/22/2010 |
| | | Soln. No.: 00079 |
| | | File Date: 02/22/2010 |

| | | |
|-----------------------|--------|--------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWJ S3-0363 |
| Solution Control Lot: | 09E066 | Expires: 05/22/2011 |
| | | Bottle No.: 0192 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:23S | | |
| Control 1 EC | 0.101% | 12:23S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 12:23S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:24S | | |
| Control 2 EC | 0.098% | 12:25S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 12:25S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:26S | | |
| Control 3 EC | 0.098% | 12:26S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 12:26S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:27S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 02/22/2010