

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00694	Calib. Date: 09/15/2010
Certification File No.:	00695	Calib. No.: 00016
Linearity File No.:	00696	Cert. Date: 09/15/2010
Solution File No.:	00756	Cert. No.: 00012
Sequential File No.:	00756	Lin. Date: 09/15/2010
		Lin. No.: 00012
		Soln. Date: 02/08/2011
		Soln. No.: 00098
		File Date: 02/08/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0194

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:20S		
Control 1 EC	0.101%	14:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22S		
Control 2 EC	0.097%	14:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24S		
Control 3 EC	0.098%	14:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. Thomas Tarr

Badge No.: 427

Date: 02/08/2011