

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00635	Calib. Date:	03/31/2010
Certification File No.:	00636	Cert. Date:	03/31/2010
Linearity File No.:	00637	Lin. Date:	03/31/2010
Solution File No.:	00642	Soln. Date:	04/29/2010
Sequential File No.:	00642	File Date:	04/29/2010

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	05/22/2011
Solution Control Lot:	09E066			Bottle No.:	0187

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:05D		
Control 1 EC	0.101%	14:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06D		
Control 2 EC	0.098%	14:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07D		
Control 3 EC	0.098%	14:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 04/29/2010