

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00767	Calib. Date: 03/24/2011
Certification File No.:	00768	Calib. No.: 00019
Linearity File No.:	00769	Cert. Date: 03/24/2011
Solution File No.:	00790	Cert. No.: 00013
Sequential File No.:	00790	Lin. Date: 03/24/2011
		Lin. No.: 00013
		Soln. Date: 05/19/2011
		Soln. No.: 00102
		File Date: 05/19/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 08/04/2012
Solution Control Lot:	10H081		Bottle No.: 0184

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:26D		
Control 1 EC	0.099%	11:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 2 EC	0.099%	11:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.099%	11:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____



Badge No.: 427

Date: 05/19/2011