

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00527	Calib. Date: 04/28/2009
Certification File No.:	00528	Calib. No.: 00013
Linearity File No.:	00529	Cert. Date: 04/28/2009
Solution File No.:	00534	Lin. Date: 04/28/2009
Sequential File No.:	00534	Lin. No.: 00009
		Soln. Date: 05/27/2009
		Soln. No.: 00068
		File Date: 05/27/2009

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0364
Control Solution %:	0.100%		Expires: 10/03/2009
Solution Control Lot:	07J048		Bottle No.: 0540

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:09D		
Control 1 EC	0.101%	13:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10D		
Control 2 EC	0.100%	13:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11D		
Control 3 EC	0.100%	13:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature probe serial number DDWAP2-215 *over*

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. Thomas Tarr

Badge No.: 427

Date: 05/27/2009