

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00527	Calib. Date: 04/28/2009
Certification File No.:	00528	Calib. No.: 00013
Linearity File No.:	00529	Cert. Date: 04/28/2009
Solution File No.:	00544	Cert. No.: 00009
Sequential File No.:	00544	Lin. Date: 04/28/2009
		Lin. No.: 00009
		Soln. Date: 07/23/2009
		Soln. No.: 00070
		File Date: 07/23/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0535

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:50D		
Control 1 EC	0.101%	12:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51D		
Control 2 EC	0.100%	12:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		
Control 3 EC	0.100%	12:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe serial number DDWAP 2-215 ~~001~~

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. J. Tarr

Badge No.: 427

Date: 07/23/2009