

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00767	Calib. Date: 03/24/2011
Certification File No.:	00768	Calib. No.: 00019
Linearity File No.:	00769	Cert. Date: 03/24/2011
Solution File No.:	00822	Cert. No.: 00013
Sequential File No.:	00822	Lin. Date: 03/24/2011
		Lin. No.: 00013
		Soln. Date: 08/11/2011
		Soln. No.: 00105
		File Date: 08/11/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0364
Solution Control Lot:	10H081	Expires: 08/04/2012
		Bottle No.: 0188

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:16D		
Control 1 EC	0.097%	13:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17D		
Control 2 EC	0.097%	13:18D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:19D		
Control 3 EC	0.098%	13:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. M. Tarr

Badge No.: 427

Date: 08/11/2011