

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00527	Calib. Date: 04/28/2009
Certification File No.:	00528	Calib. No.: 00013
Linearity File No.:	00529	Cert. Date: 04/28/2009
Solution File No.:	00573	Cert. No.: 00009
Sequential File No.:	00573	Lin. Date: 04/28/2009
		Lin. No.: 00009
		Soln. Date: 09/15/2009
		Soln. No.: 00073
		File Date: 09/15/2009

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	09E066	Expires: 05/22/2011
		Bottle No.: 0191

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:38D		
Control 1 EC	0.100%	13:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		
Control 2 EC	0.100%	13:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:41D		
Control 3 EC	0.099%	13:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____



Badge No.: 427

Date: 09/15/2009