

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE
Calibration File No.: 00767 Calib. Date: 03/24/2011 Calib. No.: 00019
Certification File No.: 00768 Cert. Date: 03/24/2011 Cert. No.: 00013
Linearity File No.: 00769 Lin. Date: 03/24/2011 Lin. No.: 00013
Solution File No.: 00830 Soln. Date: 09/08/2011 Soln. No.: 00106
Sequential File No.: 00830 File Date: 09/08/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWJ S3-0364
Control Solution %: 0.100% Expires: 08/04/2012
Solution Control Lot: 10H081 Bottle No.: 0199

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:50D		
Control 1 EC	0.097%	10:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51D		
Control 2 EC	0.098%	10:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:53D		
Control 3 EC	0.098%	10:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____



Badge No.: 427

Date: 09/08/2011