

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00841	Calib. Date:	10/26/2011
Certification File No.:	00842	Cert. Date:	10/26/2011
Linearity File No.:	00843	Lin. Date:	10/26/2011
Solution File No.:	00864	Soln. Date:	12/29/2011
Sequential File No.:	00864	File Date:	12/29/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	10H081	Expires:	08/04/2012
		Bottle No.:	0178

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:51S		
Control 1 EC	0.098%	11:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		
Control 2 EC	0.097%	11:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54S		
Control 3 EC	0.098%	11:54S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: 

Badge No.: 427

Date: 12/29/2011