

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00415	Calib. Date: 05/30/2008
Certification File No.:	00416	Cert. No.: 00010
Linearity File No.:	00417	Cert. Date: 05/30/2008
Solution File No.:	00466	Lin. Date: 05/30/2008
Sequential File No.:	00466	Lin. No.: 00006
		Soln. Date: 10/07/2008
		Soln. No.: 00057
		File Date: 10/07/2008
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWJ S3-0363
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 0546

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:14D		
Control 1 EC	0.099%	21:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15D		
Control 2 EC	0.099%	21:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16D		
Control 3 EC	0.099%	21:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe serial number DDWAP2-216 *KAT*

Changed By:

Last Name: TARR	First Name: THOMAS	MI: S
Signature: <u><i>Sgt A. Dan</i></u>	Badge No.: 427	Date: 10/07/2008