

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00481	Calib. Date: 11/25/2008
Certification File No.:	00482	Cert. Date: 11/25/2008
Linearity File No.:	00483	Lin. Date: 11/25/2008
Solution File No.:	00494	Soln. Date: 01/22/2009
Sequential File No.:	00494	File Date: 01/22/2009

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 10/03/2009
Solution Control Lot:	07J048		Bottle No.: 0534

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:26S		
Control 1 EC	0.102%	09:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27S		
Control 2 EC	0.099%	09:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29S		
Control 3 EC	0.099%	09:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature probe serial number DDWAP2-216*

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: \_\_\_\_\_

*Sgt. T. Tarr*

Badge No.: 427

Date: 01/22/2009