

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0382
Location:	WEST WINDSOR POLICE	
Calibration File No.:	00415	Calib. Date: 05/30/2008
Certification File No.:	00416	Calib. No.: 00010
Linearity File No.:	00417	Cert. Date: 05/30/2008
Solution File No.:	00480	Cert. No.: 00006
Sequential File No.:	00480	Lin. Date: 05/30/2008
		Lin. No.: 00006
		Soln. Date: 11/24/2008
		Soln. No.: 00059
		File Date: 11/24/2008

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWJ S3-0363
Control Solution %:	0.100%		Expires: 10/03/2009
Solution Control Lot:	07J048		Bottle No.: 0532

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.101%	14:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56S		
Control 2 EC	0.100%	14:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 3 EC	0.100%	14:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature probe serial number DDWAP2-216

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Sgt. T. Tarr

Badge No.: 427

Date: 11/24/2008