Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARWF-0382
Location: WEST WINDSOR POLICE

Calibration File No.:	00415	Calib. Date:	05/30/2008	Calib. No.:	00010
Certification File No.:	00416	Cert. Date:	05/30/2008	Cert. No.:	00006
Linearity File No.:	00417	Lin. Date:	05/30/2008	Lin. No.:	00006
Solution File No.:	00477	Soln. Date:	11/06/2008	Soln. No.:	00058
Sequential File No.:	00477	File Date:	11/06/2008		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	10/03/2009
Solution Control Lot:	07J048			Bottle No.:	0544

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:30S		
Control 1 EC	0.100%	11:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 2 EC	0.099%	11:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		
Control 3 EC	0.099%	11:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature probe serial number DDWAP2-216

Changed By:
Last Name: TARR First Name: THOMAS MI: S

Signature: Sgt. T. Tarr #427 Badge No.: 427
Date: 11/06/2008