

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00481	Calib. Date:	11/25/2008
Certification File No.:	00482	Cert. Date:	11/25/2008
Linearity File No.:	00483	Lin. Date:	11/25/2008
Solution File No.:	00511	Soln. Date:	03/18/2009
Sequential File No.:	00511	File Date:	03/18/2009
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	07J048	Expires:	10/03/2009
		Bottle No.:	0531

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:58S		
Control 1 EC	0.100%	12:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		
Control 2 EC	0.098%	13:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:01S		
Control 3 EC	0.098%	13:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

*Temperature probe serial number DDWAP2-216*

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: \_\_\_\_\_

*Soft. A. Davis #427*

Badge No.: 427

Date: 03/18/2009