

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE	Calib. No.:	00010
Calibration File No.:	00415	Cert. No.:	00006
Certification File No.:	00416	Lin. No.:	00006
Linearity File No.:	00417	Soln. No.:	00054
Solution File No.:	00444	Calib. Date:	05/30/2008
Sequential File No.:	00444	Cert. Date:	05/30/2008
		Lin. Date:	05/30/2008
		Soln. Date:	08/12/2008
		File Date:	08/12/2008
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWJ S3-0363
Solution Control Lot:	07J048	Expires:	10/03/2009
		Bottle No.:	0537

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:37D		
Control 1 EC	0.099%	07:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39D		
Control 2 EC	0.100%	07:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40D		
Control 3 EC	0.099%	07:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number DDWAP2-216 *[Signature]*

**Changed By:**

Last Name: TARR

First Name: THOMAS

MI: S

Signature: *[Signature]*

Badge No.: 539

Date: 08/12/2008