

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0382
Location:	WEST WINDSOR POLICE		
Calibration File No.:	00415	Calib. Date:	05/30/2008
Certification File No.:	00416	Cert. Date:	05/30/2008
Linearity File No.:	00417	Lin. Date:	05/30/2008
Solution File No.:	00457	Soln. Date:	09/09/2008
Sequential File No.:	00457	File Date:	09/09/2008

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWJ S3-0363
Control Solution %:	0.100%			Expires:	10/03/2009
Solution Control Lot:	07J048			Bottle No.:	0533

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:53D		
Control 1 EC	0.101%	19:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:54D		
Control 2 EC	0.100%	19:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:56D		
Control 3 EC	0.100%	19:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial number DDWAP2-216

Changed By:

Last Name: TARR

First Name: THOMAS

MI: S

Signature: _____

Badge No.: 539

Date: 09/09/2008